



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/766,295	01/27/2004	Guy E. Horne JR.	02579-P0027B	9340
24126	7590	05/04/2007		
ST. ONGE STEWARD JOHNSTON & REENS, LLC			EXAMINER	
986 BEDFORD STREET			KASZTEJNA, MATTHEW JOHN	
STAMFORD, CT 06905-5619				
			ART UNIT	PAPER NUMBER
			3739	
			MAIL DATE	DELIVERY MODE
			05/04/2007	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No.	Applicant(s)	
	10/766,295	HORNE ET AL.	
	Examiner	Art Unit	
	Matthew J. Kasztejna	3739	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 15 February 2007.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-7 and 10-13 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-7 and 10-13 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 14 May 2004 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date <u>1/17/07</u> . | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Notice of Amendment

In response to the amendment filed on February 15, 2007, amended claims 1 and 13 are acknowledged. The current rejections of the claims are *withdrawn*. The following new grounds of rejection are set forth:

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 1-3, 6-7 and 10-13 are rejected under 35 U.S.C. 102(e) as being anticipated by U.S. Patent No. 6,540,669 to Abe et al.

In regard to claims 1-3, 6-7 and 10-13, Abe et al. disclose an endoscope insertion shaft comprising: a tubular member 21 having an axis and including at least one aperture 25 for increasing the flexibility thereof; and a sheath comprising at least the following layer: a braided layer 22, a laminating layer 33, a wear layer 34, wherein the braided layer jackets the tubular member; and a barrier layer 231 comprising a polyester wrap disposed between the tubular member and the braided layer and jacketing the tubular member (see Figs. 3-5 and Col. 6, Lines 22-30).

Claim Rejections - 35 USC § 103

Art Unit: 3739

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 4-5 are rejected under 35 U.S.C. 103(a) as being unpatentable over U.S. Patent No. 6,540,669 to Abe et al. in view of U.S. Patent No. 6,749,560 to Konstorum et al.

In regard to claims 4-5, Abe et al. disclose an endoscope insertion shaft comprising: a tubular member 21 having an axis and including at least one aperture 25 for increasing the flexibility thereof; and a sheath comprising at least the following layer: a braided layer 22, a laminating layer 33, a wear layer 34, wherein the braided layer jackets the tubular member; and a barrier layer 231 comprising a polyester wrap disposed between the tubular member and the braided layer and jacketing the tubular member (see Figs. 3-5 and Col. 6, Lines 22-30) but are silent with respect to wherein the first set of apertures comprises at least one elongated aperture having an axis oriented at an angle to the axis of the tubular member and wherein the angle is in the range from zero to ninety degrees. Konstorum et al. disclose an analogous endoscope 10 having a shaft 14 comprised of a frame 26 and a cover 32 (see Figure 1). Frame 26 comprises a tube 40 that is preferably formed from a shape memory alloy material, such as Nitinol (see col. 3, lines 10-25). Tube 40 has slots 46 along at least part of its length where the pattern of slots can be varied and sections of slot patterns are provided, to vary the flexibility of the tube 40 (see col. 3, lines 29-51). The slots 46 provide the tube

Art Unit: 3739

40 with increased flexibility along the length of the first section 52 of the tube 40 (see col. 4, lines 1-5). Figure 3 shows that the slots 46 are positioned along a line parallel to the axis of the tube 40 and have an axis oriented at an angle to the axis of the tubular member, where the angle is in the range from zero to ninety degrees. Figure 3 also shows that the slots 46 are circumferentially positioned on the tube 40. It would have been obvious for one of ordinary skill in the art at the time the invention was made to provide the endoscope of Abe et al. with various patterns of apertures in order to vary the flexibility of the tube to a desired stiffness as taught by Konstorum et al.

Response to Arguments

Applicant's arguments with respect to claims 1-7 and 10-13 have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of

Art Unit: 3739


the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Matthew J. Kasztejna whose telephone number is (571) 272-6086. The examiner can normally be reached on Mon-Fri, 8:30-6:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda C.M. Dvorak can be reached on (571) 272-4764. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

MJK *MJK*
4/17/07



LINDA C. M. DVORAK
SUPERVISORY PATENT EXAMINER
GROUP 3700